

**Konformitätserklärung
Declaration of Conformity**

Wir

We

**B. Braun Melsungen AG
Carl-Braun-Straße 1
34212 Melsungen
Deutschland/Germany**erklären in eigener Verantwortung,
dass das/die Produkt/e**Introcan®,
Introcan® Certo,
Introcan®-W,
Introcan®-W Certo,
Introcan®-W With In-stopper**

Venenverweilkanülen

(Artikelnummern siehe Anlage I)

mit den Anforderungen der folgenden Richtlinie
übereinstimmt/übereinstimmenRichtlinie 93/42/EWG des Rates vom 14. Juni
1993über Medizinprodukte
geändert durch Richtlinie 2007/47/EG**Konformitätsbewertungsverfahren**
nach Anhang II (ausgenommen Abschnitt 4)
der oben genannten Richtlinie**Klassifizierung**
gemäß Anhang IX der
oben genannten Richtlinie
Klasse IIa**Benannte Stelle**
TÜV SÜD Product Service GmbH
Ridlerstraße 65
80339 München
Deutschland
Kennnummer 0123**Datum der ersten CE-Kennzeichnung**
1994-12**Gültig bis**
2024-05-26hereby declare in our own responsibility
that the product/s**Introcan®,
Introcan® Certo,
Introcan®-W,
Introcan®-W Certo,
Introcan®-W With In-stopper**

I. V. cannulae

(article numbers see attachment I)

is/are in compliance with the following directive

Council Directive 93/42/EEC of 14th June
1993concerning Medical Devices
amended by Directive 2007/47/EC**Conformity Assessment Procedure**
according to annex II (excluding section 4)
of the Council Directive named above**Classification**
according to annex IX of the
Council Directive named above
Class IIa**Notified Body**
TÜV SÜD Product Service GmbH
Ridlerstraße 65
80339 München
Germany
Identification number 0123**Date of first CE-marking**
1994-12**Valid until**
2024-05-26

Anlage I / Attachment I

Art.-Nr. / Art. No.	Produktname / Product name	Klasse / Class
4050122	Introcan®	Ila
4055764	Introcan® Certo	Ila
4251300	Introcan® Certo	Ila
4251318	Introcan® Certo	Ila
4251326	Introcan® Certo	Ila
4251334	Introcan® Certo	Ila
4251342	Introcan® Certo	Ila
4251350	Introcan® Certo	Ila
4251369	Introcan® Certo	Ila
4252071B	Introcan®	Ila
4252098B	Introcan®	Ila
4252110	Introcan®	Ila
4252110B	Introcan®	Ila
4252110C	Introcan®	Ila
4252136	Introcan®	Ila
4252136B	Introcan®	Ila
4252160B	Introcan®	Ila
4252217B	Introcan®	Ila
4252322B	Introcan®	Ila
4253302	Introcan®-W Certo	Ila
4253310	Introcan®-W Certo	Ila
4253329	Introcan®-W Certo	Ila
4253337	Introcan®-W Certo	Ila
4253345	Introcan®-W Certo	Ila
4253353	Introcan®-W Certo	Ila
4253361	Introcan®-W Certo	Ila
4254074B	Introcan®-W	Ila
4254074C	Introcan®-W	Ila
4254090	Introcan®-W	Ila
4254090B	Introcan®-W	Ila
4254112B	Introcan®-W	Ila
4254139B	Introcan®-W	Ila
4254171B	Introcan®-W	Ila
4254210B	Introcan®-W	Ila
4254325B	Introcan®-W	Ila
4258583	Introcan®-W With In-stopper	Ila

Art.-Nr. / Art. No.	Produktname / Product name	Klasse / Class
4258584	Introcan®-W With In-stopper	Ila
4258585	Introcan®-W With In-stopper	Ila
4258586	Introcan®-W With In-stopper	Ila

Amendment Information

Version	Description of the changes
26	Update validity

Title: Declaration of Conformity - 120-001 - Introcan Initiator: Caroline ? Herbst

This document is signed electronically in compliance with the B. Braun electronic signature policies and procedures by following persons:

UserName: Herbst, Caroline (fuehcade)
Title: Administrator Regulatory Affairs
Date: Thursday, 05 March 2020, 08:45 W. Europe Daylight Time
Meaning: Document signed as Author
=====

UserName: Brand, Thomas (brantode)
Title: HC-QM-DE08 Vice President QM for non-active Medical Devices
Date: Saturday, 07 March 2020, 21:53 W. Europe Daylight Time
Meaning: Approve Document
=====

UserName: Seidel, Stefan (seidstde)
Title: HC-RA-DE08E - Head of Regulatory Affairs CoE IV Systems
Date: Wednesday, 11 March 2020, 13:52 W. Europe Daylight Time
Meaning: Approve Document
=====